

A masterpiece – the MONO torque ratchet

 Made from a solid billet of high-quality titanium alloy



 Connects with other implant systems with the use of the short MONO insertion device



· Easy to clean

MONO circlip

· Easy assembly

Mount the MONO circlip over the finger flange of the MONO instrument.

· Additional safety

When MONO instruments are used these are additionally secured by the MONO circlip.




The MONO circlip can be autoclaved and is reusable.

The MONO circlip must be replaced if its function starts to be compromised or signs of wear appear (cracks, brittleness).

The MONO circlip is to be used only in combination with the Thommen torque ratchet and the instruments listed below.¹





MONO circlip, Viton				
	2.03.170Q2 ²			0
MONO torque ratchet, Titanio	um alloy			
Length				
110.0 mm	3.03.160			2 1000 Lot 5001
MONO insertion device, Stair	nless steel/PEEK			
		short ³	long	
Length		15.4 mm	25.8 mm	
		3.03.162	3.03.163	

			0	/p==./
MONO insertion	device for r	etentive anchor.	Stainless	steel/PEEK

Length	Ø	
26.0 mm	8.5 mm	3.03.169



MONO screwdriver, 4-lobe, Stainless steel/PEEK

	3.03.165	3.03.166	3.03.167
Shaft length	5.1 mm	10.3 mm	16.3 mm
Length	14.5 mm	22.2 mm	28.2 mm
	extra short '	short	long



The MONO circlip cannot be used with the MONO screwdriver extra short, as it does not have a PEEK finger rest.

Q2 = 2 pieces per package

The MONO insertion device, short, is compatible with instruments that are provided with a dental latch.