











ABUTMENT/VIS	COUPLES DE SERRAGE EN Ncm				
	10	15	20	25	30
Capuchon de cicatrisation façonneur gingival 	PF 3.0				
	PF 3.5				
	PF 4.0				
	PF 4.5				
	PF 5.0				
	PF 6.0				
Vis d'abutment 		PF 3.0			
		PF 3.5			
				PF 4.0	
				PF 4.5	
				PF 5.0	
Dynamic vis d'abutment 		PF 3.5			
				PF 4.0	
				PF 4.5	
				PF 5.0	
Vis d'abutment en combinaison avec l'abutment Art 		PF 3.5			
			PF 4.0		
			PF 4.5		
			PF 5.0		
Abutment VARIOmulti Abutment Novaloc® 		PF 3.5			
				PF 4.0	
				PF 4.5	
				PF 5.0	
Coiffe de protection VARIOmulti 	PF 3.5				
	PF 4.0				
	PF 4.5				
	PF 5.0				
	PF 6.0				
Vis occlusale (VARIOmulti, VARIO 17°, barre) Vis de fermeture occlusal, pour barre CAD/CAM 		PF 3.5			
		PF 4.0			
		PF 4.5			
		PF 5.0			
Ancrage sphérique 				PF 4.0	
				PF 4.5	
				PF 5.0	
Abutment ZEST® LOCATOR® 			PF 3.5		
					PF 4.0
					PF 4.5
					PF 5.0
Abutment ZEST® LOCATOR® et col pour VARIOmulti 			PF 4.0		
			PF 4.5		
			PF 5.0		
			PF 6.0		

Pour la pose définitive de la superstructure en bouche, il convient d'utiliser des vis d'abutment neuves.